

Alcotest 9510 Message Codes

- # MESSAGE**
- 0 SUCCESS
 - 1 TEST ABORTED
 - 2 TEST REPEATED
 - 3 CAL GAS SUPPLY
 - 4 ADJUST ERROR
 - 5 BLANK CHECK INCORRECT
 - 6 ALCOHOL IN AMBIENT AIR
 - 7 MOUTH ALCOHOL
 - 8 INTERFERENT DETECTED
 - 9 OUT OF MEASURING RANGE
 - 10 TESTS OUTSIDE TOLERANCE
 - 11 ALCOHOL CONCENTRATION OUT OF RANGE
 - 12 ALCOHOL CONCENTRATION NOT STABLE
 - 13 BLOWING NOT ALLOWED
 - 14 READY TO BLOW EXPIRED
 - 15 BLOWING TIME TOO SHORT
 - 17 MINIMUM VOLUME NOT ACHIEVED
 - 21 CAL STANDARD FAILED
 - 23 COMMUNICATION ERROR
 - 24 QUICK RESET
 - 28 PURGING ERROR
 - 34 INHALATION DETECTED
 - 37 RTC ERROR
 - 38 HARDWARE ERROR
 - 40 CHECK AIRWAY
 - 41 DATE INPUT TIMEOUT
 - 42 BATTERY VOLTAGE ERROR
 - 43 FLASH DELAY ERROR
 - 44 OPERATOR CANCELLED TEST
 - 45 DATALOGGING ERROR
 - 46 REFUSAL
 - 49 OPERATOR TIMEOUT
 - 50 PURGE ERROR 2
 - 51 PRINTER ERROR
 - 56 TEST TERMINATED
 - 57 MEASURE PROCEDURE INTERRUPTED
 - 61 EC AGING COMPENSATION ERROR
 - 66 NO TOLERANCE AGREEMENT
 - 67 ATN ERROR
 - 68 DIAGNOSTIC CHECK FAILED
 - 69 DATE DIFFERENCE ERROR
 - 70 DRYGAS CHECK ERROR